



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

VibrAlign, Inc.
530-G Southlake Blvd.
Richmond, VA 23236

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

L2355

Certificate Number


ANAB Approval

Certificate Valid Through: 04/22/2022
Version No. 004 Issued: 03/04/2020



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

VibrAlign, Inc.
530-G Southlake Blvd.
Richmond, VA 23236
Peter Eccleston
804-379-2250

CALIBRATION

Valid to: April 22, 2022

Certificate Number: L2355

Length – Dimensional Metrology

Table with 4 columns: Parameter/Equipment, Range, Expanded Uncertainty of Measurement (+/-), Reference Standard, Method, and/or Equipment. Rows include Laser and Inclinator.

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

Notes:

- 1. L = Length in millimeters.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. L2355.

Signature of R. D. ...
Vice President

